

# Fast test for assessing the susceptibility of *Mycobacterium tuberculosis* to isoniazid and rifampin by real-time PCR

Maria Gisele Gonçalves<sup>1</sup>, Lucila Okuyama Fukasawa<sup>1</sup>, Rosangela Siqueira Oliveira<sup>2</sup>,  
Maristela Marques Salgado<sup>1</sup>, Lee H Harrison<sup>3</sup>, Kathleen A Shutt<sup>3</sup>, Claudio Tavares Sacchi<sup>1/+</sup>

<sup>1</sup>Centro de Imunologia <sup>2</sup>Centro de Bacteriologia, Instituto Adolfo Lutz, São Paulo, SP, Brasil <sup>3</sup>Infectious Diseases Epidemiology Research Unit, University of Pittsburgh, Graduate School of Public Health, School of Medicine, Pittsburgh, Pennsylvania, USA

*Mycobacterium tuberculosis* is the bacterium that causes tuberculosis (TB), a leading cause of death from infectious disease worldwide. Rapid diagnosis of resistant strains is important for the control of TB. Real-time polymerase chain reaction (RT-PCR) assays may detect all of the mutations that occur in the *M. tuberculosis* 81-bp core region of the *rpoB* gene, which is responsible for resistance to rifampin (RIF) and codon 315 of the *katG* gene and the *inhA* ribosomal binding site, which are responsible for isoniazid (INH). The goal of this study was to assess the performance of RT-PCR compared to traditional culture-based methods for determining the drug susceptibility of *M. tuberculosis*. BACTEC™ MGIT™ 960 was used as the gold standard method for phenotypic drug susceptibility testing. Susceptibilities to INH and RIF were also determined by genotyping of *katG*, *inhA* and *rpoB* genes. RT-PCR based on molecular beacon probes was used to detect specific point mutations associated with resistance. The sensitivities of RT-PCR in detecting INH resistance using *katG* and *inhA* targets individually were 55% and 25%, respectively and 73% when combined. The sensitivity of the RT-PCR assay in detecting RIF resistance was 99%. The median time to complete the RT-PCR assay was three-four hours. The specificities for tests were both 100%. Our results confirm that RT-PCR can detect INH and RIF resistance in less than four hours with high sensitivity.

Key words: *Mycobacterium tuberculosis* - RT-PCR - drug resistance - INH - RIF

*Mycobacterium tuberculosis* is the bacterium that causes tuberculosis (TB), a leading cause of death from infectious disease worldwide. In 2008, approximately two million deaths and nine million new cases of TB were reported and 5% of these were resistant to anti-TB drugs (WHO 2009a, b). In Brazil, TB was the fourth leading cause of death from infectious disease in 2008 and the first for acquired immune deficiency syndrome patients (MS/SVS 2010), with an incidence of 48 per 100,000 inhabitants (SESSP/SVE 2006, MS/SVS 2009). From the 11,900 isolates processed in our laboratory during the last three years (2007-2009), 15% exhibited resistance to at least one drug and 5.5% were multidrug-resistant (MDR).

Rifampin (RIF) resistance is an excellent marker for MDR TB. Approximately 95% of all RIF-resistant strains contain mutations localised in an 81-bp core region of the *rpoB* gene (Telenti et al. 1993, Riska et al. 2000) and all mutations that occur in that region result in RIF resistance (Telenti et al. 1993, El-Hajj et al. 2001). In contrast to RIF, 75% of isoniazid (INH)-resistant strains have mutations either in codon 315 of the *katG* gene or in the *inhA* ribosomal binding site accompanied by mutations in the *ahpC* gene and/or the *oxyR* intergenic region (Piatek et al. 2000, Rossetti et al. 2002). In Brazil, similar percent-

ages of mutations in the *rpoB* and *katG* genes have been described (Valim et al. 2000, Rossetti et al. 2002).

Rapid diagnosis of resistant strains is important for the control of TB. The isolation and subsequent susceptibility testing of *M. tuberculosis* remain the gold standard, but this process takes many weeks to complete (Wada et al. 2004, MS/SVS/DVE 2008). The real-time polymerase chain reaction (RT-PCR) assay is faster and may detect all mutations that occur in the *M. tuberculosis* *rpoB* 81-bp core region, the *katG* 315 codon and the *inhA* ribosomal binding site (Piatek et al. 2000, El-Hajj et al. 2001, Marín et al. 2004, Ruiz et al. 2004, Wada et al. 2004, Yesilkaya et al. 2006, Boehme et al. 2010). In this study, we investigated an RT-PCR assay based on molecular beacon probes to determine susceptibility and/or resistance of *M. tuberculosis* isolates to INH and RIF in comparison with the phenotypic gold standard method that, in our study, was considered to be the automatised BACTEC MGIT 960 susceptibility assay (Becton Dickinson. Available from: [finddiagnostics.org/export/sites/default/resource-entre/find\\_documentation/pdfs/mgit\\_manual\\_nov\\_2007.pdf](http://finddiagnostics.org/export/sites/default/resource-entre/find_documentation/pdfs/mgit_manual_nov_2007.pdf)). The goal was to assess whether RT-PCR is able to produce sensitive, specific and faster results than the traditional method for routine laboratory drug-susceptibility testing.

## SUBJECTS, MATERIALS AND METHODS

*M. tuberculosis* isolates - We selected the first 988 isolates of *M. tuberculosis* received in our laboratory between October 30 2008-March 13 2009 for phenotypic drug-susceptibility testing from TB patients living in greater São Paulo, Brazil. *M. tuberculosis* lysates were prepared by transferring colonies from a Löwenstein-Jensen slant into a 2 mL screw-cap tube containing

Financial support: CNPq (573830/2008-8), FIC/Global Infectious Diseases Research Training Program, NIH, University of Pittsburgh (D43TW006592)

+ Corresponding author: [ctsacchi@gmail.com](mailto:ctsacchi@gmail.com)

Received 9 November 2011

Accepted 9 July 2012

1 mL of H<sub>2</sub>O. The suspension was boiled for 20 min and frozen at -70°C overnight. The thawed suspension was then centrifuged for 10 min at 10,000 g and the supernatant was diluted 1/10 in H<sub>2</sub>O. The lysate suspensions were kept at -20°C until used.

**Automated BD Bactec™ MGIT™ 960 test** - We used the BACTEC™ MGIT™ 960 system (Becton, Dickinson and Company, Maryland, USA) as the gold standard method for phenotypic drug susceptibility testing, following the manufacturer's instructions. The INH and RIF-susceptible *M. tuberculosis* strain H37Rv (ATCC 27294) was used as a wild type control for susceptibility testing, for RT-PCR assays and for *rpoB*, *katG* and *inhA* DNA sequencing analysis.

**PCR amplification and DNA sequencing** - The amplification conditions and nucleotide sequences of the primers used to amplify the 555-bp segment of the *katG* gene containing codons 313-318 (GenBank access X68081 - *katG* 315), a 248-bp segment of the *inhA* gene (GenBank access Z79701) and a 189-bp segment of the 81-bp core of the *rpoB* gene (GenBank access L27989) were previously described (Telenti et al. 1997, Piatek et al. 1998, 2000, El-Hajj et al. 2001, van Doorn et al. 2003, Wada et al. 2004) (Table I). DNA amplification was verified on 2% agarose gels. The amplified PCR products were purified using a PureLink™ PCR Purification Kit (Invitrogen Corporation, Carlsbad, CA, USA). DNA sequencing was performed using the Big Dye Terminator Cycle Sequencing Kit (Applied Biosystems, Foster City, CA, USA) with an ABI Prism 3130xl DNA Sequencer (Applied Biosystems). The selection of *M. tuberculosis* isolates for DNA amplification and sequencing was based on the phenotypic susceptibility testing results for INH and RIF. All *M. tuberculosis* isolates that were phenotypically resistant to INH (120) or RIF (74) were submitted to gene sequencing for *katG* and *inhA* or *rpoB*, respectively (Supplementary data). For comparison purposes, additional numbers of *M. tuberculosis* INH and RIF-sensitive isolates were also submitted for *katG* and *inhA* and *rpoB* gene sequencing, approximately 10% of the 868 (n = 90) and 914 (n = 95), respectively (Supplementary data).

**RT-PCR** - All 988 *M. tuberculosis* isolates were tested for *katG*, *inhA* and *rpoB* using single-target RT-PCR assays. The sensitivity and specificity of the RT-PCR assays were calculated using the phenotypic resistance test results as the gold standard. To test for INH resistance, two sets of primers and probes were used to hybridise to two different amplicon targets (Table I). One molecular beacon probe was complementary to codons 313-318 of the *katG* gene and the other was complementary to the *inhA* ribosomal binding site. To test for RIF resistance, one set of primers and five probes (SW143, SW89, SW111, SW112 and SW182) were used to hybridise to a single *rpoB* amplicon such that their probe sequences spanned the entire core region, with overlapping sequences of one to three nucleotides (Table I). The RT-PCR probes were molecular beacons designated to hybridise to targets that did not contain mutations associated with antibiotic resistance. Drug resistance was indicated by the absence of a characteristic increasing fluorescent signal during PCR amplification. All assays were performed in duplicate and

in sealed wells on a 96-well microtitre plate (Roche Diagnostics, Indianapolis, Ind). Reactions were performed in a Roche LightCycler 480 II (Roche Diagnostics, Indianapolis, Ind). Each reaction mixture contained 5 µL of lysed suspension diluted 1/10 in water, 2 µL of each primer, 2 µL of probe and 12.5 µL of 2X TaqMan® Universal master mix (Applied Biosystems). PCR-certified water (Roche Diagnostics, Indianapolis, Ind) was added to bring the reaction volume to 25 µL. PCR mixtures were first incubated for 2 min at 50°C and 10 min at 95°C and then for 45 cycles of 30 s at 95°C, annealing at 1 min at 50°C (55°C for the *inhA* reaction) and 30 s at 60°C.

All reactions for RIF testing were initially performed as a single-target RT-PCR assay with one set of primers and one FAM-labelled probe per reaction tube for all 988 isolates. After all probes had been individually tested, all phenotypically RIF-resistant and the same ~10% additional RIF-sensitive isolates mentioned above were re-evaluated by dual-target RT-PCR assay. We prepared two sets of dual-target RT-PCR assays: (i) the HEX-labelled probe SW182 plus the FAM-labelled probe SW112 and (ii) the HEX-labelled probe SW89 plus the FAM-labelled probe SW111. Each combination was used in a different RT-PCR assay. Probe SW143 was not used for dual-target assays because we did not identify any mutation in that region based on a single-target RT-PCR assay or by DNA sequencing.

C<sub>T</sub> values from 18-39 were considered positive, meaning that no mutation was present on the gene fragment being tested, therefore indicating susceptibility. C<sub>T</sub> values equal to zero or equal to or greater than 40 were considered negative, indicating the presence of a drug resistance mutation. C<sub>T</sub> values between 1-17 were not found among the tested isolates.

**Ethics** - This study was approved by the ethical review board of the São Paulo University (CONEP/265/08).

## RESULTS

**Phenotypic determination of resistance** - Of the 988 isolates studied, 120 (12%) and 74 (7%) were phenotypically INH or RIF-resistant, respectively, and 65 (7%) were resistant to both drugs. Sixty-five (88%) of the 74 RIF-resistant isolates were also resistant to INH. The overall sensitivity to both drugs determined by the phenotypic test was 87% (859/988) [95% confidence interval (CI): 85, 89].

**Mutation frequencies** - We characterised mutations in the *katG*, *inhA* and *rpoB* gene loci. Among INH-resistant isolates, mutations of the *katG* gene (55%; 66/120) were more frequent than in the *inhA* gene (25%; 30/120); only eight (7%; 8/120) presented mutations on both genes. No mutations were identified among the 90 INH-sensitive isolates submitted to DNA sequencing.

**katG** - Among INH-resistant isolates, all mutations were at codon 315. At this codon, the substitution from AGC to ACC, leading to the amino acid change serine (S) to threonine (T), was observed in 59 (91%) isolates. In addition, single-nucleotide polymorphisms from AGC (S) to AAC (N), AGA (R), CGC (R) and ACG (T) were seen in two (3%), two (3%), one (1.5%) and one (1.5%) isolates, respectively (Table II).

TABLE I  
Primers and probes for gene amplification, sequencing and real-time polymerase chain reaction (RT-PCR)

Gene target	Primers and probes	Name	Sequence (5'-3')	T <sub>m</sub> (°C)	Final concentration (µM)	Reference
katG	Amplification	katGf1	GCAGATGGGGCTGATCTACG	60	0.4	Wada et al. (2004)
		katGrl	AACTCGTCGGCCCAATTCCTC			
	Sequencing	katGf2	GGGCTTGGGCTGGAAGAG	50	0.16	van Doorn et al. (2003)
		katGr2	ACAGGATCTCGAGGAAACTGTTGT			
	RT-PCR primers	katGf3	CGTCGGGGTCACACTTTCGGTAAGA	50	0.6	Piatek et al. (2000)
		katGr3	TTGTCCCAATTCGTCTGGGGGTTCGT		0.6	
	RT-PCR probe	katGpb	FAM-CCGAGGCACCAGCGGCATCGACCTCGG-BHQ1	0.3		
inhA	Amplification and sequencing	TB92	CCTCGCTGCCAGAAAGGGA	58	0.4 and 0.16	Talenti et al. (1997)
		TB93	ATCCCCGGTTTCCTCCGGT			
rpoB	RT-PCR primers	inhAf	GTGGACATAACCGATTTCG	55	0.3	Piatek et al. (2000)
		inhAr	CTCCGGTAACAGGACTGAACGGG		0.3	
	RT-PCR probe	inhAPb	FAM-CGAGGCCGACAACCTATCGTCTCCCTCG-BHQ1		0.1	
		rpoBf1	GGAGGCGATCACACCCGACAGCGTT	60	0.4 and 0.16	El Hajji et al. (2001)
rpoB	Amplification and sequencing	rpoBr1	ACCTCCAGCCCCGGCACGCTCACGT			
		rpoBf2	GGCCGGTGGTCCGCCG	50	0.3	Piatek et al. (1998)
	rpoBr2	ACGTGACAGACCCCGCGGC		0.3		
	RT-PCR probes	SW143	FAM-CGAGCTCAGCTGGTGGGCTCG		0.3	
		SW89	FAM or HEX-GCTACGGAGCCAATTCATGGACACAGTAGC		0.3	
		SW111	FAM-CCAGCCGACAGCGGGTGTTCGTGG		0.3	
SW112		FAM-CCACGCTTGTGGGTCAACCCCGTGG		0.3		
SW182	FAM or HEX-CCTGGGCCGACTGTCCGGGCTGCCAGG		0.3			

T<sub>m</sub>: temperature of melting.

*inhA* - Among INH-resistant isolates, mutations at the ribosomal binding site at the -15 or -16 positions were found with substitutions from CGT-CAT (87%) and CTC-ATC (13%) (Table II).

*rpoB* - Among RIF-resistant (74) and RIF-sensitive isolates (95), mutations at the 81-bp core fragment region of the *rpoB* gene were investigated; 75 resistant and 94 sensitive isolates were identified by sequencing (Supplementary data). The most frequently mutated codon was 531 (67%): TCG-TTG (64%; 48/74) and TCG-TGG (3%; 2/75), followed by codon 526 (Table II) (23%; 17/75), codon 516 (GAC-GTC) (8%; 6/75) and codons 522 (TCG-TTG) and 533 (CTG-CCG) (1%; 1/75 each). Codon 526 (23%; 17/75) was polymorphic, with five different substitutions from CAC: GAC (47%; 8/17), TAC (35%; 6/17), TGC, CTC and TGC (6%; 1/17 each) (Table II). All mutations detected by DNA sequencing were located in four of the five different probe complementary regions of the 81-bp core region of the *rpoB* gene (SW89, SW111, SW112 and SW182). No mutations were detected in the SW143 probe complementary region.

*Single-target RT-PCR assay - INH* - Of the 988 isolates analysed, 91% (900/988) were sensitive and 9% (88/988) were resistant as determined by RT-PCR (Supplementary data). The overall sensitivities of the assay in detecting INH resistance using *katG* and *inhA* targets individually were 55% (66/120) and 25% (30/120), respectively (95% CI: 46-64 and 18-34, respectively). When the results of both targets were combined, the overall sensitivity was 73% (88/120) (95% CI: 64-81). The specificities for the tests were both 100% (95% CI: 97-100). All of the mutations identified by sequencing the *katG* and *inhA* gene probe complementary regions were detected by RT-PCR

assays. Therefore, the specificity and sensitivity in detecting the point mutations in both genes were 100%. No discrepant results were obtained between duplicates.

*RIF* - Of the 988 isolates analysed, 92.4% (913/988) were sensitive and 7.6% (75/988) were resistant as determined by RT-PCR (Supplementary data). The overall sensitivity of the RT-PCR assay in detecting RIF resistance using all five different probes was 99% (73/74) (95% CI: 93, 100) when compared to the phenotypic assay. One of the RIF phenotypically resistant isolates was RIF-sensitive by RT-PCR assay. The nucleotide sequence analysis of this isolate revealed that a mutation at codon 522 (TCG-TTG), covered by probe SW111, was not detected by RT-PCR assay. We found a discrepancy between the phenotypic and genotypic results (RT-PCR and DNA sequencing) for RIF resistance for two *M. tuberculosis* isolates: one at codon 516 (GAC-TAC) and the other at codon 526 (CAC-CTC), using probes SW89 and SW112, respectively. No discrepant results were obtained between duplicates.

*Dual-target RT-PCR assay* - The combination of two probes for each of the dual-target RT-PCR assays for RIF was based on the prevalence of mutations at the 81-bp core region of the *rpoB* gene. We excluded probe SW143 from both combinations because there were no mutations present on the DNA region complementary to probe SW143. All 74 phenotypically RIF-resistant isolates and 95 (10%) of phenotypically RIF-sensitive isolates were used to evaluate the two dual-target RT-PCR assays (SW182 + SW112 and SW89 + SW111). There was 100% correlation between the results of single and dual-target RT-PCR assays. No discrepant results were obtained between duplicates.

TABLE II  
Gene mutation found associated with anti-tuberculosis drug resistance in São Paulo isolates

Gene	Isolates with mutation	DNA position		Base changed to	Isolates (n)	Change (%)	Codon (%)
		Number	Bases				
<i>rpoB</i>	75	516	GAC	GTC	5	7	8
			TAC	1	1	-	
		522	TCG	TTG	1	1	1
			CAC	CGC	1	1	23
		526	TGC	1	1	-	
			CTC	1	1	-	
			GAC	8	12	-	
			TAC	6	8	-	
			TTG	48	64	67	
		531	TCG	TTG	2	3	-
				TGG	1	1	1
533	CTG	CCG	1	1	1		
		CCG	1	1	1		
<i>katG</i>	65	315	AGC	ACC	59	91	100
				AAC	2	3	-
				AGA	2	3	-
				CGC	1	1.5	-
				ACG	1	1.5	-
				ACG	1	1.5	-
<i>inhA</i>	30	-15	G	A	26	87	100
		-16	C	A	4	13	-

## DISCUSSION

Because of the slow growth of *M. tuberculosis*, long delays in the diagnosis of drug-resistant TB occur when conventional culture-based susceptibility assays are used (the time required is, on average, 2-4 weeks). More rapid approaches that use genetic analysis for the detection of mutations associated with drug resistance have been described (Telenti et al. 1993, Kapur et al. 1994, Torres et al. 2000, 2003, El-Hajj et al. 2001, Garcia de Viedma et al. 2002, Garcia de Viedma 2003, van Doorn et al. 2003, Marín et al. 2004, Wada et al. 2004, Yesilkaya et al. 2006, Boehme et al. 2010) and the RT-PCR assay has been proposed as an alternative approach to detect drug-resistant organisms (Kapur et al. 1994, Piatek et al. 2000, Garcia de Viedma et al. 2002, Marín et al. 2004, Ruiz et al. 2004, Kocagoz et al. 2005, Boehme et al. 2010). In this study, we assessed reported RT-PCR assays to predict the drug resistance of *M. tuberculosis* isolates.

The RT-PCR results and DNA sequencing analysis obtained in this study were consistent for all *katG* and *inhA* mutations, except for one *rpoB* mutation. In this isolate, probe SW111 was not able to detect a mutation at codon 522 (TCG-TTG). We are unable to explain this discrepancy despite numerous repetitions of MGIT 960, DNA sequencing and RT-PCR assays. The mutation missed by probe SW111 is close to the 3' end of its sequence, only three nucleotides before the molecular beacon hairpin arm. It is possible that the binding capacity of probe SW111 is affected when mismatches are too close to that area. We also identified discrepancy between phenotypic and genotypic results of the RIF resistance of two *M. tuberculosis* isolates. DNA sequencing predicted that RIF resistance of these two isolates was consistent after three rounds of DNA sequencing and RT-PCR testing, but they were phenotypically RIF-sensitive, even after repetitions. According to our data, RT-PCR and MGIT failed to detect 1% (1/74) and 2% (2/95) of RIF-resistant isolates, respectively.

If we consider that any mutation on the 81-bp core region of the *rpoB* gene confers RIF resistance, these two cases suggest that sometimes RT-PCR may be more accurate than the phenotypic assay. Similar discrepancies have been described, such as when a mutation at codon 533 was compatible with a clinically RIF-resistant case, but was overlooked by the BACTEC method (Riska et al. 2000).

All 75 (100%) RIF-resistant isolates obtained by genetic sequencing in this study had mutations in the *rpoB* gene. This contrasts with previous data showing that RIF-resistant isolates in Brazil exhibited *rpoB* mutation frequencies varying from 80-89% (Valim et al. 2000, de Miranda et al. 2001, Clemente et al. 2008). Previous data on *katG* reported that mutations occur in approximately 60-85% of INH-resistant isolates (Rossetti et al. 2002, Cardoso et al. 2004, Hofling et al. 2005, Clemente et al. 2008) but, in our study, mutation of the *katG* gene was seen in 55% (65/118) of the INH-resistant isolates. The reasons for these discrepancies are not clear, but may be related to different methods used to perform the phenotypic susceptibility testing, differences in year and geographical site of *M. tuberculosis* isolation and the use of different sizes of sample collections.

RT-PCR cannot completely replace culture-based susceptibility tests because conventional testing can detect resistance due both to genetic mechanisms that are not yet understood and to a larger number of drugs. There are a number of limitations to the widespread use of PCR-based techniques; resistance to anti-TB drugs may involve changes in multiple genes and at multiple possible locations within a gene. However, the fact that the majority of the INH and RIF resistance of *M. tuberculosis* isolates are due to point mutations on the *katG*, *inhA* and *rpoB* genes, respectively, molecular approaches such as RT-PCR are promising.

Our results confirm that RIF-resistant *M. tuberculosis* can be detected in less than 4 h using a very specific and sensitive assay. Moreover, the assay is a powerful tool because it is simple to perform and readily automatable for high-throughput screening and the results enable faster diagnosis and immediate decision-making to determine appropriate drug therapy.

The rapid detection of *M. tuberculosis* resistance continues to be an important public health matter. Commercial assays, such as GeneXpert (Cepheid), which detects *M. tuberculosis* and resistance to RIF, demonstrated good performance with clinical specimens (Helb et al. 2010, Lawn & Nicol 2011, Zeka et al. 2011). In our study, the use of RT-PCR to determine *M. tuberculosis* resistance to RIF was also very promising and its application in clinical samples will be explored in future studies.

The cost of in-house molecular designs could be equivalent to, or even less expensive than, the costs of phenotypic antibiograms if the new liquid-culture media are used. However, cost, speed and professional expertise are not the only factors to consider. The molecular analysis of resistance provides additional information and essential data that are not available if only phenotypic antibiograms are obtained. Genotypic analysis also specifies the resistance mutation involved, which helps to identify strains with high and/or broad levels of resistance (van Soolingen et al. 2000, Garcia de Viedma 2003). The optimal approach for drug susceptibility testing of MTB will likely involve both methodologies.

## ACKNOWLEDGEMENTS

To Marcílio Figueiredo Lemos, for technical assistance on DNA sequencing, to Fábio Oliveira Latrilha, for general technical assistance, and to Glavur Matte and Maria Helena Matte, for assistance with the initial study design.

## REFERENCES

- Boehme CC, Nabeta P, Hillemann D, Nicol MP, Shenai S, Krapp F, Allen J, Tahirli R, Blakemore R, Rustomjee R, Milovic A, Jones M, O'Brien SM, Persing DH, Ruesch-Gerdes S, Gotuzzo E, Rodrigues C, Alland D, Perkins MD 2010. Rapid molecular detection of tuberculosis and rifampin resistance. *N Engl J Med* 363: 1005-1015.
- Cardoso RF, Cooksey RC, Morlok GP, Barco P, Cecon L, Forestiero F, Leite CQ, Sato DN, de Lourdes Shikama M, Mamizuka EM, Hirata MH 2004. Screening and characterization of mutations in isoniazid-resistant *Mycobacterium tuberculosis* isolates obtained in Brazil. *Antimicrob Agents Chemother* 48: 3373-3381.
- Clemente WT, Lima SSS, Palaci M, Silva MSN, Rodrigues VFS, Costa ERD, Possuelo L, Cafrune PI, Ribeiro FK, Gomes HM, Serufo JC 2008. Phenotypic and genotypic characterization of

- drug-resistant *Mycobacterium tuberculosis* strains. *Diag Microbiol Infect Dis* 62: 199-204.
- de Miranda SS, Kritski AL, Filliol I, Mabilat C, Panteix G, Drouet E 2001. Mutations in the *rpoB* gene of rifampicin-resistant *Mycobacterium tuberculosis* strains isolated in Brazil and France. *Mem Inst Oswaldo Cruz* 96: 247-250.
- El-Hajj HH, Marras SAE, Tyagi S, Kramer FR, Alland D 2001. Detection of rifampin resistance in *Mycobacterium tuberculosis* in a single tube with molecular beacons. *J Clin Microbiol* 39: 4131-4137.
- Garcia de Viedma D 2003. Rapid detection of resistance in *Mycobacterium tuberculosis*: a review discussing molecular approaches. *Clin Microbiol Infect* 9: 349-359.
- Garcia de Viedma D, Infantes MSD, Lasala F, Chaves F, Alcalá L, Bouza E 2002. New real-time PCR able to detect in a single tube multiple rifampin resistance mutations and high-level isoniazid resistance mutations in *Mycobacterium tuberculosis*. *J Clin Microbiol* 40: 988-995.
- Helb D, Jones M, Story E, Boehme C, Wallace E, Ho K 2010. Rapid detection of *Mycobacterium tuberculosis* and rifampin-resistance using on-demand, near-patient technology. *J Clin Microbiol* 1: 229-237.
- Hofling CC, Pavan EM, Giampaglia CMS, Ferrazoli L, Aily DCB, de Albuquerque DM, Ramos MC 2005. Prevalence of *katG* Ser315 substitution and *rpoB* mutations in isoniazid-resistant *Mycobacterium tuberculosis* isolates from Brazil. *Int J Tuberc Lung Dis* 9: 87-89.
- Kapur V, Li LL, Iordanescu S, Hamrick MR, Wanger A, Kreiswirth BN, Musser JM 1994. Characterization by automatic DNA sequencing of mutations in the gene (*rpoB*) encoding the RNA polymerase beta subunit in rifampin-resistant *Mycobacterium tuberculosis* strains from New York city and Texas. *J Clin Microbiol* 32: 1095-1098.
- Kocagoz T, Saribas Z, Alp A 2005. Rapid determination of rifampin resistance in clinical isolates of *Mycobacterium tuberculosis* by real-time PCR. *J Clin Microbiol* 43: 6015-6019.
- Lawn SD, Nicol MP 2011. Xpert® MTB/RIF assay: development, evaluation and implementation of a new rapid molecular diagnostic for tuberculosis and rifampicin resistance. Review. *Future Microbiol* 6: 1067-1082.
- Marín M, Viedma DG, Ruiz-Serrano MJ, Bouza E 2004. Rapid direct detection of multiple rifampin and isoniazid resistance mutations in *Mycobacterium tuberculosis* in respiratory samples by real-time PCR. *Antimicrob Agents Chemother* 48: 4293-4300.
- MS/SVS - Ministério da Saúde/Secretaria de Vigilância em Saúde 2009. Informe eletrônico de tuberculose. Boletim epidemiológico 2. Available from: [portal.saude.gov.br/portal/arquivos/pdf/boletim\\_tb\\_julho09.pdf](http://portal.saude.gov.br/portal/arquivos/pdf/boletim_tb_julho09.pdf).
- MS/SVS - Ministério da Saúde/Secretaria de Vigilância em Saúde 2010. Programa Nacional de Controle da Tuberculose 2010. Manual de recomendações para o controle da tuberculose no Brasil. Available from: [portal.saude.gov.br/portal/arquivos/pdf/manual-derecomendacoescontroletbno.pdf](http://portal.saude.gov.br/portal/arquivos/pdf/manual-derecomendacoescontroletbno.pdf).
- MS/SVS/DVE - Ministério da Saúde/Secretaria de Vigilância em Saúde/Departamento de Vigilância Epidemiológica 2008 *Manual nacional de vigilância laboratorial da tuberculose e outras micobactérias*, 1st ed., MS/SVS/DVE, Brasília, p. 323-348.
- Piatek AS, Telenti A, Murray MR, El-Hajj H, Jacobs Jr WR, Kramer FR, Alland D 2000. Genotypic analysis of *Mycobacterium tuberculosis* in two distinct populations using molecular beacons: implications for rapid susceptibility testing. *Antimicrob Agents Chemother* 44: 103-110.
- Piatek AS, Tyagi S, Pol AC, Telenti A, Miller LP, Kramer FR, Alland D 1998. Molecular beacon sequence analysis for detecting drug resistance in *Mycobacterium tuberculosis*. *Nat Biotechnol* 16: 359-363.
- Riska PF, Jacobs Jr WR, Alland D 2000. Molecular determinants of drug resistance in tuberculosis. *Int J Tuberc Lung Dis* 4 (Suppl.): S4-S10.
- Rossetti MLR, Valim ARM, Silva MSN, Rodrigues VS 2002. Tuberculose resistente: revisão molecular. *Rev Saude Publica* 36: 525-532.
- Ruiz M, Torres MJ, Llanos AC, Arroyo A, Palomares JC, Aznar J 2004. Direct detection of rifampin and isoniazid-resistant *Mycobacterium tuberculosis* in auramine-rhodamine-positive sputum specimens by real-time PCR. *J Clin Microbiol* 42: 1585-1589.
- SESSP/SVE - Secretaria de Estado da Saúde de São Paulo/Centro de Vigilância Epidemiológica 2006. Tuberculose no estado de São Paulo. Indicadores de morbimortalidade e indicadores de desempenho. Boletim Epidemiológico Paulista 3 (Suppl. 4). Available from: [ftp://ftp.cve.saude.sp.gov.br/doctec/outros/BEPA\\_TB-SUPLE4V3.pdf](http://ftp.cve.saude.sp.gov.br/doctec/outros/BEPA_TB-SUPLE4V3.pdf).
- Telenti A, Honore N, Bernasconi C, March J, Ortega A, Heym B, Takiff HE, Cole ST 1997. Genotypic assessment of isoniazid and rifampin resistance in *Mycobacterium tuberculosis*: a blind study at reference laboratorial level. *J Clin Microbiol* 35: 719-723.
- Telenti A, Imboden P, Marchesi F, Matter L, Schopfer K, Bodmer T, Lowrie D, Cole S, Colston MJ, Cole S 1993. Detection of rifampicin resistance mutations in *Mycobacterium tuberculosis*. *Lancet* 341: 647-650.
- Torres JM, Criado A, Ruiz M, Llanos AC, Palomares JC, Aznar J 2003. Improved real-time PCR for rapid detection of rifampin and isoniazid resistance in *Mycobacterium tuberculosis* clinical isolates. *Diag Microbiol Infect Dis* 45: 207-212.
- Torres MJ, Criado A, Palomares JC, Aznar J 2000. Use of real-time PCR and fluorimetry for rapid detection of rifampin resistance-associated mutations in *Mycobacterium tuberculosis*. *J Clin Microbiol* 38: 3194-3199.
- Valim ARM, Rossetti MLR, Ribeiro MO, Zaha A 2000. Mutations in the *rpoB* gene of multidrug-resistant *Mycobacterium tuberculosis* isolates from Brasil. *J Clin Microbiol* 38: 3119-3122.
- van Doorn HR, Class EC, Templeton KE, van der Zanden AGM, Vije AK, de Jong MD, Dankert J, Kuijper EJ 2003. Detection of a point mutation associated with high-level isoniazid resistance in *Mycobacterium tuberculosis* by real-time PCR technology with 3'-minor groove binder-DNA probes. *J Clin Microbiol* 41: 4630-4635.
- van Soolingen D, Hass PEW, van Door HR, Kuijper E, Rinder H, Borgdorff MW 2000. Mutations at amino acid position 315 of the *katG* gene area associated with high-level resistance to isoniazid, other drug resistance, and successful transmission of *Mycobacterium tuberculosis* in the Netherlands. Concise communication. *J Infect Dis* 182: 1788-1790.
- Wada T, Maeda S, Tamaru A, Imai S, Hase A, Kobayashi K 2004. Dual-probe assay for rapid detection of drug-resistant *Mycobacterium tuberculosis* by real-time PCR. *J Clin Microbiol* 42: 5277-5285.
- WHO - World Health Organization 2009a. Global tuberculosis control report. Available from: [who.int/tb/publications/2009/tbfactsheet\\_2009update\\_one\\_page.pdf](http://who.int/tb/publications/2009/tbfactsheet_2009update_one_page.pdf).
- WHO - World Health Organization 2009b. Global tuberculosis control. A short update to the 2009 report. Available from: [who.int/tb/publications/global\\_report/2009/update/tb\\_9.pdf](http://who.int/tb/publications/global_report/2009/update/tb_9.pdf).
- Yesilkaya H, Meacci F, Niemann S, Hillemann D, Rüh-Gerdes S, Barer MR, Andrew PW, Oggioni MR, Long Drug Study Group 2006. Evaluation of molecular-beacon, TaqMan and fluorescence resonance energy transfer probes for detection of antibiotic resistance-conferring single nucleotide polymorphisms in mixed *Mycobacterium tuberculosis* DNA extracts. *J Clin Microbiol* 44: 3826-3829.
- Zeka ZN, Tasbakan S, Cavusoglu C 2011. Evaluation of the GeneXpert MTB/RIF assay for rapid diagnosis of tuberculosis and detection of rifampin resistance in pulmonary and extrapulmonary specimens. *J Clin Microbiol* 49: 4138-4414.

Susceptibility and resistance to isoniazid (INH),  
rifampin (RIF) or both in 988 isolates obtained  
by phenotypic and molecular methods  
[real-time polymerase chain reaction (RT-PCR)]

Methods	Susceptible (S) or resistant (R)	INH	RIF	INH + RIF
		n (%)	n (%)	n (%)
Phenotypic	S	868 (88)	914 (93)	859 (87)
	R	120 (12)	74 (7)	65 (7)
RT-PCR	S	900 (91)	913 (92.4)	878 (89)
	R	88 (9)	75 (7.6)	110 (11)